RayBio® Human bFGF ELISA Kit

User Manual (Revised Mar 1, 2012)

RayBio[®] Human bFGF ELISA Kit Protocol

(Cat#: ELH-bFGF-001)



We Provide You With Excellent Protein Array System And Service

Tel:(Toll Free)1-888-494-8555 or 770-729-2992; Fax:770-206-2393; web: www.raybiotech.com Email: info@raybiotech.com



RayBio® Human bFGF ELISA Kit Protocol

TABLE OF CONTENTS

I.	Introduction	
II.	Reagents	2
III.	Storage	3
IV.	Additional Materials Required	3
V.	Reagent Preparation	4
VI.	Assay Procedure	6
VII.	Assay Procedure Summary	7
VIII.	Calculation of Results	
A.	Typical Data	8
B.	Sensitivity	8
C.	Recovery	8
D.	Linearity	9
E.	Reproducibility	9
IX.	Specificity	9
X.	References	10
XI.	Troubleshooting Guide	11

I. INTRODUCTION

bFGF (basic fibroblast growth factor) is found in almost all tissues of mesodermal and neuroectodermal origin and also in tumors derived from these tissues. Endothelial cells produce large amounts of this factor. Some bFGF is associated with the extracellular matrix of the subendothelial cells. bFGF is an 18 kDa protein with a length of 155 amino acids and an isoelectric point of 9.6. It does not contain disulfide bonds and is not glycosylated. bFGF stimulates the growth of fibroblasts, myoblasts, osteoblasts, neuronal cells, endothelial cells, keratinocytes, chondrocytes, and many other cell types. bFGF has been shown to be a promoting or inhibitory modulator of cellular differentiation also for other cell types. bFGF is not only a mitogen for chondrocytes but also inhibits their terminal differentiation.

The RayBio® Human bFGF ELISA (Enzyme-Linked Immunosorbent Assay) kit is an in vitro enzyme-linked immunosorbent assay for the quantitative measurement of human bFGF in serum, plasma, cell culture supernatants and urine. This assay employs an antibody specific for human bFGF coated on a 96-well plate. Standards and samples are pipetted into the wells and bFGF present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti-human bFGF antibody is added. After washing away unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of bFGF bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm.

II. REAGENTS

1. bFGF Microplate (Item A): 96 wells (12 strips x 8 wells) coated with anti-human bFGF.

- 2. Wash Buffer Concentrate (20x) (Item B): 25 ml of 20x concentrated solution.
- 3. Standards (Item C): 2 vials, recombinant human bFGF.
- 4. Assay Diluent A (Item D): 30 ml, 0.09% sodium azide as preservative. For Standard/Sample (serum/plasma) diluent.
- 5. Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For Standard/Sample (cell culture medium/urine) diluent.
- 6. Detection Antibody bFGF (Item F): 2 vial of biotinylated anti-human bFGF (each vial is enough to assay half microplate).
- 7. HRP-Streptavidin Concentrate (Item G): 200 µl of 120x concentrated HRP-conjugated streptavidin.
- 8. TMB One-Step Substrate Reagent (Item H): 12 ml of 3,3',5,5'-tetramethylbenzidine (TMB) in buffered solution.
- 9. Stop Solution (Item I): 8 ml of 0.2 M sulfuric acid.

III. STORAGE

May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard (recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after reconstitution. Opened Microplate Wells or reagents may be store for up to 1 month at 2° to 8°C. Return unused wells to the pouch containing desiccant pack, reseal along entire edge.

Note: the kit can be used within one year if the whole kit is stored at -20 °C. Avoid repeated freeze-thaw cycles.

IV. ADDITIONAL MATERIALS REQUIRED

- 1 Microplate reader capable of measuring absorbance at 450 nm.
- 2 Precision pipettes to deliver 2 µl to 1 ml volumes.
- 3 Adjustable 1-25 ml pipettes for reagent preparation.
- 4 100 ml and 1 liter graduated cylinders.
- 5 Absorbent paper.
- 6 Distilled or deionized water.


- 7 Log-log graph paper or computer and software for ELISA data analysis.
- 8 Tubes to prepare standard or sample dilutions.

V. REAGENT PREPARATION

- 1. Bring all reagents and samples to room temperature (18 25°C) before use.
- 2. Sample dilution: If your samples need to be diluted, Assay Diluent A (Item D) should be used for dilution of serum/plasma samples. 1x Assay Diluent B (Item E) should be used for dilution of culture supernantants and urine.

Suggested dilution for normal serum/plasma: 2 fold*.

- *Please note that levels of the target protein may vary between different specimens. Optimal dilution factors for each sample must be determined by the investigator.
- 3. Assay Diluent B should be diluted 5-fold with deionized or distilled water.
- 4. Preparation of standard: **Briefly spin the vial of Item C** and then add 400 μl Assay Diluent A (for serum/plasma samples) or 1x Assay Diluent B (for cell culture medium and urine) into Item C vial to prepare a 100 ng/ml standard. **Dissolve the powder thoroughly by a gentle mix.** Add 100 μl bFGF standard from the vial of Item C, into a tube with 900 μl Assay Diluent A or 1x Assay Diluent B to prepare a 10,000 pg/ml stock standard solution. Pipette 300 μl Assay Diluent A or 1x Assay Diluent B into each tube. Use the stock standard solution to produce a dilution series (shown below). Mix each tube thoroughly before the next transfer. Assay Diluent A or 1x Assay Diluent B serves as the zero standard (0 pg/ml).



- 5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1x Wash Buffer.
- 6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100 μl of 1x Assay Diluent B into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days). The detection antibody concentrate should be diluted 65-fold with 1x Assay Diluent B and used in step 4 of Part VI Assay Procedure.
- 7. Briefly spin the HRP-Streptavidin concentrate vial (Item G) and pipette up and down to mix gently before use. HRP-Streptavidin concentrate should be diluted 120-fold with 1x Assay Diluent B.

For example: Briefly spin the vial (Item G) and pipette up and down to mix gently. Add 100 μ l of HRP-Streptavidin concentrate into a tube with 12 ml 1x Assay Diluent B tocprepare a final 120 fold diluted HRP-Streptavidin solution. Mix well.

VI. ASSAY PROCEDURE:

- 1. Bring all reagents and samples to room temperature (18 25°C) before use. It is recommended that all standards and samples be run at least in duplicate.
- 2. Add 100 µl of each standard (see Reagent Preparation step 2) and sample into appropriate wells. Cover well and incubate for 2.5 hours at room temperature or over night at 4°C with gentle shaking.
- 3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with Wash Buffer (300 µl) using a multi-channel Pipette or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
- 4. Add 100 μl of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well. Incubate for 1 hour at room temperature with gentle shaking.
- 5. Discard the solution. Repeat the wash as in step 3.
- 6. Add 100 µl of prepared Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle shaking.
- 7. Discard the solution. Repeat the wash as in step 3.
- 8. Add 100 µl of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.
- 9. Add 50 µl of Stop Solution (Item I) to each well. Read at 450 nm immediately.

VII. ASSAY PROCEDURE SUMMARY

1. Prepare all reagents, samples and standards as instructed.

Ŋ

2. Add 100 μl standard or sample to each well. Incubate 2.5 hours at room temperature or over night at 4°C.

 \iint

3. Add 100 µl prepared biotin antibody to each well. Incubate 1 hour at room temperature.

 \int

4. Add 100 μl prepared Streptavidin solution. Incubate 45 minutes at room temperature.

 \int

5. Add 100 µl TMB One-Step Substrate Reagent to each well. Incubate 30 minutes at room temperature.

 $\hat{\mathbb{I}}$


6. Add 50 µl Stop Solution to each well. Read at 450 nm immediately.


VIII. CALCULATION OF RESULTS

Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average zero standard optical density. Plot the standard curve on log-log graph paper or using Sigma plot software, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.

A. TYPICAL DATA

These standard curves are for demonstration only. A standard curve must be run with each assay.





B. SENSITIVITY

The minimum detectable dose of bFGF is typically less than 50 pg/ml.

C. RECOVERY

Recovery was determined by spiking various levels of human bFGF into human serum, plasma and cell culture media. Mean recoveries are as follows:

Sample Type	Average % Recovery	Range (%)
Serum	95.46	84-104
Plasma	92.64	82-103
Cell culture media	93.48	83-103

D. LINEARITY

Sample Type		Serum	Plasma	Cell Culture Media
1:2	Average % of Expected Range (%)	91 85-103	89 84-103	92 84-103
1:4	Average % of Expected Range (%)	93 83-104	92 83-104	90 84-105

E. REPRODUCIBILITY

Intra-Assay: CV<10% Inter-Assay: CV<12%

IX. SPECIFICITY

Cross Reactivity: This ELISA kit shows no cross-reactivity with any of the cytokines tested (*e.g.*, human Angiogenin, BDNF, BLC, ENA-78, FGF-4, IL-1 α , IL-1 β , IL-2, IL-3, IL-4, IL-5, IL-7, IL-8, IL-9, IL-10, IL-11, IL-12 p70, IL-12 p40, IL-13, IL-15, IL-309, IP-10, G-CSF, GM-CSF, IFN- γ , Leptin, MCP-1, MCP-2, MCP-3, MDC, MIP-1 α , MIP-1 β , MIP-1 δ , PARC, PDGF, RANTES, SCF, TARC, TGF- β , TIMP-1, TIMP-2, TNF- α , TNF- β , TPO, VEGF).

X. REFERENCES:

- 1. Baird A and Klagsbrun M. The fibroblast growth factor family. Cancer Cells 3: 239-43 (1991).
- 2. Burgess WH and Maciag T. The heparin-binding (fibroblast) growth factor family of proteins. Annual Review of Biochemistry 58: 575-606 (1989).
- 3. Bashkin P et al Basic fibroblast growth factor binds to subendothelial extracellular matrix and is released by heparitinase and heparin-like molecules. Biochemistry 28: 1737-43 (1989).

XI. TROUBLESHOOTING GUIDE

Problem	Cause	Solution
Poor standard curve	Inaccurate pipetting	1. Check pipettes
	Improper standard dilution	 Ensure a brief spin of Item C and dissolve the powder thoroughly by a gentle mix.
2. Low signal	1.Too brief incubation times	 Ensure sufficient incubation time; assay procedure step 2 may change to over night
	Inadequate reagent volumes or improper dilution	Check pipettes and ensure correct preparation
3. Large CV	Inaccurate pipetting	 Check pipettes
4. High background	Plate is insufficiently washed	 Review the manual for proper wash. If using a plate washer, check that all ports are unobstructed.
	Contaminated wash buffer	Make fresh wash buffer
5. Low sensitivity	Improper storage of the ELISA kit	1. Store your standard at<-20°C after reconstitution, others at 4 °C. Keep substrate solution protected from light
	2. Stop solution	2. Stop solution should be added to each well before measure

RayBio® ELISA kits:

Over 300 ELISA kits, custom ELISA kit choose from over 500 list visit www.raybiotech.com for details.

This product is for research use only.

