

PLD5000C Surgical X-ray Machine

This equipment must only be operated by the professional who has necessary radiation knowledge and get professional training.

This machine is used to photograph patient's esophagus, chest, gastrointestinal, abdomen, limbs etc for common radiography, Beam Limiting Device radiograph and high KV radiography.



Feature:

- 1.The console applies LCD display photograph term, the device have the mA, kV of photograph table projection chosen are controlled by procedure interlock, be separate ,not disturb.
- 2.Use the microprocessor limited-time control, to achieve the precise control of the expose time.
- 3.Select the exposure tube current at the same time to determine the focus.
- 4.This machine adopts mA, KV, s three parameters free matching work method.
- 5.The limit time of fluoroscope of this machine is controlled by program.
- 6.Expose mode of this machine use two-step hand-swicth control mod, hand swicth control circuit adopt a safe voltage control.