

Results Driven Research

Human Interleukin-1β (IL-1 β) ELISA Kit

Introduction

Interleukin-1beta (IL-1beta) has a wide spectrum of inflammatory, metabolic, haemopoietic, and immunological properties (1). IL-1 beta plays a significant role in hippocampal synaptic function (2), and is a potential genetic marker as an indicator of gastric cancer risk (3). High plasma levels of Interleukin-1 beta (IL-1b) is associated with rheumatoid and osteoarthritic joint disease (4), infectious gastroenteritis (5), neurodegeneration (6), and breast cancer (7). High gingival crevicular fluid levels of interleukin-1beta is related to type 2 diabetes (8).

Principal of the Assay

The Human IL-1beta ELISA kit is designed for detection of human IL-1beta in plasma, serum, and cell culture supernatants. This assay employs a quantitative sandwich enzyme immunoassay technique, which measures IL-1beta in less than 5 hours. A murine monoclonal antibody specific for IL-1beta has been pre-coated onto a microplate. IL-1beta in standards and samples is sandwiched by the immobilized antibody and a biotinylated polyclonal antibody specific for IL-1beta, which is recognized by a streptavidin-peroxidase conjugate. All unbound material is then washed away and a peroxidase enzyme substrate is added. The color development is stopped and the intensity of the color is measured.

Caution and Warning

- Prepare all reagents (working diluent buffer, wash buffer, standards, biotinylated-antibody, and SP conjugate) as instructed, prior to running the assay.
- Prepare all samples prior to running the assay. The dilution factors for the samples are suggested in this protocol. However, the user should determine the optimal dilution factor.
- Spin down the SP conjugate vial and the biotinylated-antibody vial before opening and using contents.
- This kit is for research use only.
- The kit should not be used beyond the expiration date.
- The Stop Solution is an acid solution.

Reagents

• **IL-1beta Microplate:** A 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against IL-1beta.

- **Sealing Tapes:** Each kit contains 3 pre-cut, pressure-sensitive sealing tapes that can be cut to fit the format of the individual assay.
- **IL-1beta Standard:** Human IL-1beta in a buffered protein base (500 pg, lyophilized).
- **Biotinylated IL-1beta Antibody** (**100x**): A 100-fold biotinylated polyclonal antibody against human IL-1beta (80 µl).
- MIX Diluent Concentrate (10x): A 10-fold concentrated buffered protein base (30 ml).
- Wash Buffer Concentrate (20x): A 20-fold concentrated buffered surfactant (30 ml, 2 bottles).
- Streptavidin-Peroxidase Conjugate (SP Conjugate): A 100-fold concentrate (80 µl).
- **Chromogen Substrate**: A ready-to-use stabilized peroxidase chromogen substrate tetramethylbenzidine (8 ml).
- Stop Solution: A 0.5 N hydrochloric acid to stop the chromogen substrate reaction (12 ml).

Storage Condition

- Store components of the kit at 2-8°C or -20°C upon arrival up to the expiration date.
- Store SP Conjugate and Biotinylated Antibody at -20^oC
- Store Microplate, Diluent Concentrate (10x), Wash Buffer, Stop Solution, and Chromogen Substrate at 2-8°C
- Opened unused microplate wells may be returned to the foil pouch with the desiccant packs. Reseal along zip-seal. May be stored for up to 1 month in a vacuum desiccator.
- Diluent (1x) may be stored for up to 1 month at 2-8°C.
- Store Standard at 2-8°C before reconstituting with Diluent and at -20°C after reconstituting with Diluent.

Other Supplies Required

- Microplate reader capable of measuring absorbance at 450 nm
- Pipettes (1-20 μl, 20-200 μl, 200-1000 μl and multiple channel)
- Deionized or distilled reagent grade water

Sample Collection, Preparation and Storage

- **Plasma:** Collect plasma using one-tenth volume of 0.1 M sodium citrate as an anticoagulant. Centrifuge samples at 2000 x g for 10 minutes and assay. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- **Serum:** Samples should be collected into a serum separator tube. After clot formation, centrifuge samples at 2000 x g for 10 minutes. Remove serum and assay. The undiluted samples can be stored at -20°C or below for up to 3 months. Avoid repeated freeze-thaw cycles.
- Cell Culture Supernatants: Centrifuge cell culture media at 2000 x g for 10 minutes to remove debris. Collect supernatants and assay. Store the remaining samples at -20°C or below. Avoid repeated freeze-thaw cycles.

Reagent Preparation

- Freshly dilute all reagents and bring all reagents to room temperature before use.
- MIX Diluent Concentrate (10x): If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute the MIX Diluent 1:10 with reagent grade water. Store for up to 1 month at 2-8°C.
- Standard Curve: Reconstitute the 500 pg of human IL-1beta Standard with 2 ml of MIX Diluent to generate a standard solution of 250 pg/ml. Allow the standard to sit for 10 minutes with gentle agitation prior to making dilutions. Prepare duplicate or triplicate standard points by serially diluting the IL-1beta standard solution (250 pg/ml) 1:2 with equal volume of MIX Diluent to produce 125, 62.5, 31.3, 15.7, 7.8, and 3.9 pg/ml solutions. MIX Diluent serves as the zero standard (0 pg/ml). Any remaining solution should be frozen at -20°C.

Standard Point	Dilution	[IL-1beta] (pg/ml)
P1	1 part Standard (250 pg/ml)	250.0
P2	1 part P1 + 1 part MIX Diluent	125.0
P3	1 part P2 + 1 part MIX Diluent	62.50
P4	1 part P3 + 1 part MIX Diluent	31.25
P5	1 part P4 + 1 part MIX Diluent	15.63
P6	1 part P5 + 1 part MIX Diluent	7.810
P7	1 part P6 + 1 part MIX Diluent	3.910
P8	MIX Diluent	0.000

- **Biotinylated IL-1beta Antibody** (**100x**): Spin down the antibody briefly and dilute the desired amount of the antibody 1:100 with MIX Diluent. Any remaining solution should be frozen at -20°C.
- Wash Buffer Concentrate (20x): If crystals have formed in the concentrate, mix gently until the crystals have completely dissolved. Dilute Wash Buffer 1:20 with reagent grade water.
- **SP Conjugate (100x):** Spin down the SP Conjugate briefly and dilute the desired amount of the conjugate 1:100 with MIX Diluent. Any remaining solution should be frozen at -20^oC.

Assay Procedure

- Prepare all reagents, working standards and samples as instructed. Bring all reagents to room temperature before use. The assay is performed at room temperature (20-30°C).
- Remove excess microplate strips from the plate frame and return them immediately to the foil pouch with desiccant inside. Reseal the pouch securely to minimize exposure to water vapor and store in a vacuum desiccator.
- Add 50 μl of Standard or sample per well. Cover wells with a sealing tape and incubate for two hours. Start the timer after the last sample addition.
- Wash five times with 200 μl of Wash Buffer manually. Invert the plate each time and decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid. If using a machine wash six times with 300 μl of Wash Buffer and then invert the plate, decant the contents; hit it 4-5 times on absorbent paper towel to completely remove the liquid.
- Add 50 µl of Biotinylated IL-1beta Antibody to each well and incubate for two hours.
- Wash the microplate as described above.

- Add 50 μl of Streptavidin-Peroxidase Conjugate per well and incubate for 30 minutes. Turn on the microplate reader and set up the program in advance.
- Wash the microplate as described above.
- Add 50 µl of Chromogen Substrate per well and incubate for about 10 minutes or until the optimal blue color density develops. Gently tap the plate to ensure thorough mixing and break the bubbles in the well with pipette tip.
- Add 50 µl of Stop Solution to each well. The color will change from blue to yellow.
- Read the absorbance on a microplate reader at a wavelength of 450 nm immediately. If wavelength correction is available, subtract readings at 570 nm from those at 450 nm to correct optical imperfections. Otherwise, read the plate at 450 nm only. Please note that some unstable black particles may be generated at high concentration points after stopping the reaction for about 10 minutes, which will reduce the readings


Data Analysis

- Calculate the mean value of the duplicate or triplicate readings for each standard and sample.
- To generate a standard curve, plot the graph using the standard concentrations on the x-axis and the corresponding mean 450 nm absorbance on the y-axis. The best-fit line can be determined by regression analysis using log-log or four-parameter logistic curve-fit.
- Determine the unknown sample concentration from the Standard Curve and multiply the value by the dilution factor.

Standard Curve

• The curve is provided for illustration only. A standard curve should be generated each time the assay is performed.

H. IL-1beta Standard Curve



Performance Characteristics

- The minimum detectable level of IL-1beta is typically ~ 3 pg/ml.
- Intra-assay and inter-assay coefficients of variation were 5.1 % and 7.5% respectively.

Linearity

	Average Percentage of Expected Value	
Sample Dilution	Plasma	Serum
No dilution	96%	95%
1:2	94%	89%
1:4	93%	88%

Cross-Reactivity

Species	% Cross Reactivity	
Canine	None	
Monkey	None	
Mouse	None	
Rat	None	
Swine	None	
Bovine	None	
Rabbit	None	

References

- (1) Palmi M et al. (2002) Mol Neurobiol. 25(2): 133-47
- (2) Lynch MA (2002) Vitam Horm. 64:185-219
- (3) Troost E et al. (2003) Can J Gastroenterol. 17 Suppl B: 8B-12B
- (4) Fan Z et al. (2004) Cytokine 28(1): 17-24
- (5) Enocksson A et al. (2004) Clin Diagn Lab Immunol. 11(2): 250-4
- (6) Allan SM et al. (2005) Nat Rev Immunol 5(8): 629-40
- (7) Mettler L et al. (2004) Clin Exp Obstet Gynecol. 31(1): 20-2
- (8) Engebretson SP et al. (2004) J Periodontol. 75(9): 1203-8

Version 3.8